

CERTIFICATE OF CONFORMANCE

Part Number: W3T101573 **Serial Number:** LAB2909
Model Number: IP-LABXT10 **Date Tested:** 3/7/2018

The above stated Model/Part Number is certified to have been manufactured and tested in accordance with our ISO 9001 Quality and all the requirements specified in Evoqua Water Technologies LLC Manufacturing Work Instruction MFG-074.

<u>TEST PROCEDURE</u>	<u>ACCEPTANCE CRITERIA</u>	<u>PASSED TEST</u>
1) Pressure Test (MFG-073)	25 psi for ≥ 0.25 hours - Leak Free	<input checked="" type="checkbox"/>
2) Rinse up Test (MFG-073)	≥ 15 MOhm-cm - ≥ 7 MOhm-cm w/200 uS salt	<input checked="" type="checkbox"/>
3) Wiring Test (MF125ST)	Electrical harness function	<input checked="" type="checkbox"/>

<u>APPLICABLE STANDARDS</u>	<u>CONFORMS</u>
1) Asbestos / Lead Free (RoHS Compliant)	<input checked="" type="checkbox"/>
2) Wetted Materials- FDA requirements for food contact applications (21 CFR)	<input checked="" type="checkbox"/>
3) This product has been certified as HALAL	<input checked="" type="checkbox"/>

The manual for the above component can be found on: www.ionpure.com
Should you need further assistance, please E-mail: ionpure@evoqua.com
We value you as our customer and will accommodate you in whatever way to provide total satisfaction.

Sincerely,



Rahoul Bhagat - ISO QA Mgmt. Rep.

Evoqua Water Technologies LLC - Ionpure Products