

We the undersigned manufacturer
Evoqua Water Technologies
10 Technology Drive
Lowell, MA, USA
Phone: (978) 441-6025 Fax: (978) 614-7058

Certify and declare under our sole responsibility that the following product(s):

<i>IP-MXM030</i>	<i>IP-MXM060</i>	<i>IP-MXM125</i>	<i>IP-MXM250</i>	<i>IP-MXM500</i>

Manufactured for Evoqua Water Technologies at:
10 Technology Drive, Lowell, MA 01851

Conforms to the following directives
2006/95/EC - Low voltage

The units conform to the essential requirements of the LV directive based on the following harmonized standard (s)

1. BS EN 60204-1: 2006 Safety of Machinery – electrical equipment of machines – Part 1: General Requirements
2. Continuity of the protective bonding circuit (Measured resistance < 160 mΩ)
3. Insulation Resistance test (> 1 MΩ)
4. Voltage test (leakage current < current limit)

The technical documentation file is located at
Evoqua Water Technologies, 10 Technology Drive, Lowell, MA, USA

The authorized representative within the community is:
Steve Willis, Evoqua Water Technologies, Global Sales Manager
PO BOX 1135, Hanslope, Milton Keynes, MK19 7ZQ, UK



Name and position of person binding the manufacturer:
Rahoul Bhagat, Quality Assurance and Document Control
11 July 2014

